

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022719\
 Data File : VU029680.D
 Acq On : 26 Feb 2019 14:31
 Operator : JC/SP
 Sample : VU0227WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Quant Time: Feb 27 02:49:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Feb 27 01:53:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	481102	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	453326	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	218746	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	155912	44.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.46%
7) Chloroethane-d5	1.68	69	139393	44.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	88.96%
11) 1,1-Dichloroethene-d2	2.27	63	233953	33.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.90%
21) 2-Butanone-d5	4.17	46	227242	105.92	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.92%
24) Chloroform-d	4.64	84	298329	52.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.36%
26) 1,2-Dichloroethane-d4	5.31	65	184200	50.88	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.76%
32) Benzene-d6	5.34	84	621072	54.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.90%
36) 1,2-Dichloropropane-d6	6.32	67	194491	56.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	113.34%
41) Toluene-d8	7.56	98	593928	51.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.04%
43) trans-1,3-Dichloropropene-	7.85	79	89632	49.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.48%
47) 2-Hexanone-d5	8.31	63	176485	92.35	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.35%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	287984	49.66	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.32%
64) 1,2-Dichlorobenzene-d4	11.86	152	240693	51.06	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.12%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU022719\
Data File : VU029680.D
Acq On : 26 Feb 2019 14:31
Operator : JC/SP
Sample : VU0227WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Feb 27 02:49:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Feb 27 01:53:00 2019
Response via : Initial Calibration

