

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022719\
 Data File : VU029703.D
 Acq On : 27 Feb 2019 14:35
 Operator : JC/SP
 Sample : K1621-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Quant Time: Feb 28 01:21:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 28 00:50:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	455023	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	439235	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	211918	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	169164	50.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	101.46%
7) Chloroethane-d5	1.68	69	142774	48.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	96.34%
11) 1,1-Dichloroethene-d2	2.27	63	229988	35.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.58%
21) 2-Butanone-d5	4.17	46	255561	125.95	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	125.95%
24) Chloroform-d	4.65	84	286421	52.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.92%
26) 1,2-Dichloroethane-d4	5.31	65	186988	54.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.22%
32) Benzene-d6	5.34	84	655147	59.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	118.56%
36) 1,2-Dichloropropane-d6	6.33	67	197413	59.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	118.74%
41) Toluene-d8	7.56	98	614231	54.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.92%
43) trans-1,3-Dichloropropene-	7.85	79	92342	52.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.70%
47) 2-Hexanone-d5	8.31	63	210173	113.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	113.50%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	297546	52.96	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	105.92%
64) 1,2-Dichlorobenzene-d4	11.86	152	248311	54.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.74%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU022719\
 Data File : VU029703.D
 Acq On : 27 Feb 2019 14:35
 Operator : JC/SP
 Sample : K1621-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Quant Time: Feb 28 01:21:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Feb 28 00:50:18 2019
 Response via : Initial Calibration

