

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU022819\
 Data File : VU029733.D
 Acq On : 28 Feb 2019 13:18
 Operator : JC/SP
 Sample : K1577-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CWJW4

Quant Time: Mar 01 00:55:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 00:18:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	259670	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	249126	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	126131	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	74969	4.42	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.40%
7) Chloroethane-d5	1.68	69	73942	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.60%
11) 1,1-Dichloroethene-d2	2.27	63	118452	3.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.00%
20) 2-Butanone-d5	4.19	46	215845	56.29	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.58%
24) Chloroform-d	4.65	84	153127	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
26) 1,2-Dichloroethane-d4	5.31	65	82809	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.34	84	333061	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
36) 1,2-Dichloropropane-d6	6.33	67	102877	5.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.60%
41) Toluene-d8	7.56	98	298344	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
43) trans-1,3-Dichloropropene-	7.85	79	38225	5.00	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.00%
46) 2-Hexanone-d5	8.31	63	211814	56.57	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.14%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	87171	5.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	127414	5.14	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.80%

Target Compounds					Qvalue
13) Acetone	2.32	43	7581	2.752 ug/L	98
34) Trichloroethene	6.18	95	10367	0.564 ug/L	94
42) Toluene	7.64	91	43665	0.585 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU022819\
Data File : VU029733.D
Acq On : 28 Feb 2019 13:18
Operator : JC/SP
Sample : K1577-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CWJW4

Quant Time: Mar 01 00:55:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 00:18:18 2019
Response via : Initial Calibration

