

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU022819\
 Data File : VU029736.D
 Acq On : 28 Feb 2019 14:29
 Operator : JC/SP
 Sample : K1577-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CWJX2

Quant Time: Mar 01 01:05:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 00:18:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	255678	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	245049	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	125736	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	75325	4.51	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.20%
7) Chloroethane-d5	1.68	69	75993	4.84	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.80%
11) 1,1-Dichloroethene-d2	2.27	63	119755	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	4.19	46	226881	60.10	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.20%
24) Chloroform-d	4.65	84	154905	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
26) 1,2-Dichloroethane-d4	5.31	65	84432	5.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.20%
32) Benzene-d6	5.34	84	337068	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.20%
36) 1,2-Dichloropropane-d6	6.33	67	103854	5.67	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.40%
41) Toluene-d8	7.56	98	310804	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
43) trans-1,3-Dichloropropene-	7.85	79	38773	5.16	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.20%
46) 2-Hexanone-d5	8.31	63	217853	59.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.30%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	88518	5.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	130158	5.26	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.20%

Target Compounds

					Qvalue
13) Acetone	2.32	43	6390	2.356 ug/L	100
22) cis-1,2-Dichloroethene	4.22	96	84744	4.483 ug/L	93
34) Trichloroethene	6.19	95	4320	0.239 ug/L	89
42) Toluene	7.64	91	90852	1.237 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU022819\
Data File : VU029736.D
Acq On : 28 Feb 2019 14:29
Operator : JC/SP
Sample : K1577-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CWJX2

Quant Time: Mar 01 01:05:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 00:18:18 2019
Response via : Initial Calibration

