

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030119\
 Data File : VU029751.D
 Acq On : 01 Mar 2019 09:50
 Operator : JC/SP
 Sample : VU0301WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK45

Quant Time: Mar 01 22:20:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 22:12:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	234411	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	226405	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	110275	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	73672	4.81	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.20%
7) Chloroethane-d5	1.68	69	70786	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.40%
11) 1,1-Dichloroethene-d2	2.27	63	114322	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	4.19	46	213215	61.60	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.20%
24) Chloroform-d	4.65	84	148874	5.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.40%
26) 1,2-Dichloroethane-d4	5.31	65	82540	5.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.40%
32) Benzene-d6	5.34	84	330645	5.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	113.80%
36) 1,2-Dichloropropane-d6	6.33	67	100330	5.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	118.60%
41) Toluene-d8	7.56	98	291925	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
43) trans-1,3-Dichloropropene-	7.85	79	35019	5.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.80%
46) 2-Hexanone-d5	8.31	63	204720	60.16	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	120.32%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	85538	5.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	112.60%
64) 1,2-Dichlorobenzene-d4	11.86	152	122535	5.65	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.00%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030119\
Data File : VU029751.D
Acq On : 01 Mar 2019 09:50
Operator : JC/SP
Sample : VU0301WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK45

Quant Time: Mar 01 22:20:41 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 22:12:01 2019
Response via : Initial Calibration

