

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030119\
 Data File : VU029766.D
 Acq On : 01 Mar 2019 16:06
 Operator : JC/SP
 Sample : K1660-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CWJY8

Quant Time: Mar 04 11:58:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 22:12:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	235368	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	226870	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	116266	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	88246	5.74	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.80%
7) Chloroethane-d5	1.68	69	68928	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.40%
11) 1,1-Dichloroethene-d2	2.27	63	120094	3.64	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.80%
20) 2-Butanone-d5	4.19	46	180630	51.97	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.94%
24) Chloroform-d	4.65	84	120483	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
26) 1,2-Dichloroethane-d4	5.31	65	72380	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
32) Benzene-d6	5.34	84	304595	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
36) 1,2-Dichloropropane-d6	6.33	67	91327	5.39	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.80%
41) Toluene-d8	7.56	98	278667	4.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.80%
43) trans-1,3-Dichloropropene-	7.85	79	33751	4.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.00%
46) 2-Hexanone-d5	8.31	63	173602	50.91	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.82%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	72191	4.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	114405	5.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Acetone	2.32	43	100090	40.088	ug/L	98
22) cis-1,2-Dichloroethene	4.22	96	24198	1.391	ug/L	98
34) Trichloroethene	6.18	95	5812	0.347	ug/L	89
42) Toluene	7.64	91	49915	0.734	ug/L	99
47) Tetrachloroethene	8.22	164	9832	0.633	ug/L	95
51) Chlorobenzene	9.12	112	18352	0.393	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030119\
Data File : VU029766.D
Acq On : 01 Mar 2019 16:06
Operator : JC/SP
Sample : K1660-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CWJY8

Quant Time: Mar 04 11:58:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 22:12:01 2019
Response via : Initial Calibration

