

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030119\
 Data File : VU029770.D
 Acq On : 01 Mar 2019 18:29
 Operator : JC/SP
 Sample : K1577-21
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Mar 04 12:03:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 01 22:12:01 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	226274	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	220690	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	108537	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	84510	5.72	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.40%
7) Chloroethane-d5	1.68	69	70067	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.27	63	118522	3.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.60%
20) 2-Butanone-d5	4.19	46	171449	51.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.62%
24) Chloroform-d	4.65	84	122785	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
26) 1,2-Dichloroethane-d4	5.31	65	69831	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
32) Benzene-d6	5.34	84	293939	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
36) 1,2-Dichloropropane-d6	6.33	67	88365	5.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.20%
41) Toluene-d8	7.56	98	273018	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
43) trans-1,3-Dichloropropene-	7.85	79	32391	4.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.60%
46) 2-Hexanone-d5	8.31	63	165144	49.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.58%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	70597	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.40%
64) 1,2-Dichlorobenzene-d4	11.86	152	110765	5.19	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030119\
Data File : VU029770.D
Acq On : 01 Mar 2019 18:29
Operator : JC/SP
Sample : K1577-21
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Mar 04 12:03:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 01 22:12:01 2019
Response via : Initial Calibration

