

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030220\
 Data File : VU036991.D
 Acq On : 02 Mar 2020 12:16
 Operator : JC/MD
 Sample : VU0302MBL01
 Misc : 5.0µl/5.0ml/100µL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK36

Quant Time: Mar 03 03:32:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 03 03:21:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	255202	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	243068	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	88443	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	87497	40.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.66%
7) Chloroethane-d5	1.93	69	82967	46.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.58%
11) 1,1-Dichloroethene-d2	2.59	63	118235	32.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.88%
21) 2-Butanone-d5	4.67	46	167930	116.82	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.82%
24) Chloroform-d	5.10	84	166328	50.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.64%
26) 1,2-Dichloroethane-d4	5.74	65	117644	53.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.24%
32) Benzene-d6	5.76	84	356365	50.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.40%
36) 1,2-Dichloropropane-d6	6.72	67	121424	53.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	106.64%
41) Toluene-d8	7.92	98	322488	49.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.90%
43) trans-1,3-Dichloropropene-	8.20	79	51458	48.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.24%
47) 2-Hexanone-d5	8.66	63	116919	105.00	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.00%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	174568	54.37	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.74%
64) 1,2-Dichlorobenzene-d4	12.21	152	99351	57.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	114.92%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030220\
Data File : VU036991.D
Acq On : 02 Mar 2020 12:16
Operator : JC/MD
Sample : VU0302MBL01
Misc : 5.0µl/5.0ml/100µl/5.0ml/MSVOA_U/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK36

Quant Time: Mar 03 03:32:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 03 03:21:45 2020
Response via : Initial Calibration

