

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030220\
 Data File : VU036993.D
 Acq On : 02 Mar 2020 13:01
 Operator : JC/MD
 Sample : L1708-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Mar 03 03:39:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 03 03:21:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	258510	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	248003	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	93460	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	88143	40.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.20%
7) Chloroethane-d5	1.93	69	84051	46.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.58%
11) 1,1-Dichloroethene-d2	2.59	63	119413	32.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.70%
21) 2-Butanone-d5	4.67	46	173217	118.95	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	118.95%
24) Chloroform-d	5.11	84	167070	50.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.80%
26) 1,2-Dichloroethane-d4	5.74	65	119347	53.20	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.40%
32) Benzene-d6	5.76	84	363597	50.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.40%
36) 1,2-Dichloropropane-d6	6.72	67	122352	52.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.32%
41) Toluene-d8	7.93	98	326915	49.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.26%
43) trans-1,3-Dichloropropene-	8.20	79	53647	49.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.36%
47) 2-Hexanone-d5	8.66	63	127525	112.25	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.25%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	176609	53.91	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.82%
64) 1,2-Dichlorobenzene-d4	12.21	152	101303	55.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.90%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030220\
Data File : VU036993.D
Acq On : 02 Mar 2020 13:01
Operator : JC/MD
Sample : L1708-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Mar 03 03:39:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 03 03:21:45 2020
Response via : Initial Calibration

