

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030220\
 Data File : VU037010.D
 Acq On : 02 Mar 2020 19:30
 Operator : JC/MD
 Sample : L1729-13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AX1

Quant Time: Mar 03 04:50:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 03 03:21:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	260499	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	244024	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	90283	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	79333	36.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.52%
7) Chloroethane-d5	1.93	69	77889	43.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.06%
11) 1,1-Dichloroethene-d2	2.59	63	109433	29.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.74%#
21) 2-Butanone-d5	4.67	46	163098	111.15	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.15%
24) Chloroform-d	5.11	84	161395	48.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.62%
26) 1,2-Dichloroethane-d4	5.74	65	113149	50.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.10%
32) Benzene-d6	5.76	84	344674	48.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.68%
36) 1,2-Dichloropropane-d6	6.72	67	115006	50.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.60%
41) Toluene-d8	7.92	98	310067	47.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.72%
43) trans-1,3-Dichloropropene-	8.20	79	48744	45.88	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.76%
47) 2-Hexanone-d5	8.66	63	114347	102.29	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	102.29%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	165776	51.43	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	102.86%
64) 1,2-Dichlorobenzene-d4	12.21	152	97768	55.40	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.80%


Target Compounds					Ovalue
13) Acetone	2.66	43	3089	2.047 ug/L	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030220\
Data File : VU037010.D
Acq On : 02 Mar 2020 19:30
Operator : JC/MD
Sample : L1729-13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AX1

Quant Time: Mar 03 04:50:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 03 03:21:45 2020
Response via : Initial Calibration





#13

Acetone

Concen: 2.047 ug/L

RT: 2.66 min Scan# 409

Delta R.T. 0.00 min

Lab File: VU037010.D

Acq: 02 Mar 2020 19:30

Instrument :

MSVOA_U

ClientSampleId :

C0AX1

Tot Ion: 43 Resp: 3089

Ion Ratio Lower Upper

43 100

58 28.1 0.0 64.6

