

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030223\
 Data File : VU053179.D
 Acq On : 02 Mar 2023 11:51
 Operator : JC/MD
 Sample : 01745-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COLFO

Quant Time: Mar 03 01:45:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021423WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 03 01:44:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.244	114	187257	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.411	117	174480	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	87064	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.598	65	133962	77.876	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 155.760%#		
7) Chloroethane-d5	1.913	69	124602	85.585	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 171.160%#		
11) 1,1-Dichloroethene-d2	2.566	63	192021	65.078	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 130.160%#		
21) 2-Butanone-d5	4.620	46	183385	203.981	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 203.980%#		
24) Chloroform-d	5.058	84	286559	94.433	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 188.860%#		
26) 1,2-Dichloroethane-d4	5.698	65	186588	96.691	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 193.380%#		
32) Benzene-d6	5.723	84	594450	95.504	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 191.000%#		
36) 1,2-Dichloropropane-d6	6.685	67	196840	96.803	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 193.600%#		
41) Toluene-d8	7.894	98	548751	95.215	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 190.420%#		
43) trans-1,3-Dichloroprop...	8.173	79	96895	93.561	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 187.120%#		
47) 2-Hexanone-d5	8.627	63	137571	195.853	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 195.850%#		
56) 1,1,2,2-Tetrachloroeth...	10.749	84	266106	98.326	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 196.660%#		
66) 1,2-Dichlorobenzene-d4	12.189	152	182896	94.495	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 189.000%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030223\
Data File : VU053179.D
Acq On : 02 Mar 2023 11:51
Operator : JC/MD
Sample : 01745-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
COLFO

Quant Time: Mar 03 01:45:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM021423WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 03 01:44:16 2023
Response via : Initial Calibration

