

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU030419\
 Data File : VU029799.D
 Acq On : 04 Mar 2019 21:32
 Operator : JC/SP
 Sample : K1661-13
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFC75

Quant Time: Mar 05 02:47:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 05 01:28:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	482246	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	464436	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	216414	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	131425	37.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.38%
7) Chloroethane-d5	1.68	69	107841	34.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	68.66%#
11) 1,1-Dichloroethene-d2	2.27	63	176304	25.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	51.04%#
21) 2-Butanone-d5	4.17	46	219139	101.90	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.90%
24) Chloroform-d	4.65	84	224570	39.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.36%
26) 1,2-Dichloroethane-d4	5.31	65	141187	38.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.82%
32) Benzene-d6	5.34	84	485677	41.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.12%
36) 1,2-Dichloropropane-d6	6.33	67	148048	42.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.22%
41) Toluene-d8	7.56	98	453175	38.00	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.00%#
43) trans-1,3-Dichloropropene-	7.85	79	66233	35.51	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.02%
47) 2-Hexanone-d5	8.31	63	191954	98.04	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	98.04%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	233628	39.33	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	78.66%
64) 1,2-Dichlorobenzene-d4	11.86	152	184123	39.48	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.96%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU030419\
Data File : VU029799.D
Acq On : 04 Mar 2019 21:32
Operator : JC/SP
Sample : K1661-13
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFC75

Quant Time: Mar 05 02:47:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 05 01:28:42 2019
Response via : Initial Calibration

