

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU030419\
 Data File : VU029790.D
 Acq On : 04 Mar 2019 18:03
 Operator : JC/SP
 Sample : K1661-01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFC70

Quant Time: Mar 05 01:52:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 05 01:28:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	489717	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	467919	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	212639	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	138832	38.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.38%
7) Chloroethane-d5	1.68	69	114237	35.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.62%
11) 1,1-Dichloroethene-d2	2.27	63	182120	25.96	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	51.92%#
21) 2-Butanone-d5	4.17	46	239668	109.75	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.75%
24) Chloroform-d	4.65	84	237551	40.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.62%
26) 1,2-Dichloroethane-d4	5.31	65	147045	39.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.80%
32) Benzene-d6	5.34	84	514715	43.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.44%
36) 1,2-Dichloropropane-d6	6.33	67	157869	44.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.14%
41) Toluene-d8	7.56	98	478661	39.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.68%#
43) trans-1,3-Dichloropropene-	7.85	79	71786	38.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.42%
47) 2-Hexanone-d5	8.31	63	211287	107.11	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.11%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	246190	41.13	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.26%
64) 1,2-Dichlorobenzene-d4	11.86	152	185860	40.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.12%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU030419\
Data File : VU029790.D
Acq On : 04 Mar 2019 18:03
Operator : JC/SP
Sample : K1661-01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFC70

Quant Time: Mar 05 01:52:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 05 01:28:42 2019
Response via : Initial Calibration

