

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU030419\
 Data File : VU029804.D
 Acq On : 04 Mar 2019 23:28
 Operator : JC/SP
 Sample : K1661-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Mar 05 03:01:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 05 01:28:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	508442	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	487887	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	237072	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	163002	43.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.50%
7) Chloroethane-d5	1.68	69	135021	40.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	81.54%
11) 1,1-Dichloroethene-d2	2.27	63	230815	31.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.38%
21) 2-Butanone-d5	4.17	46	230027	101.46	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.46%
24) Chloroform-d	4.64	84	283380	46.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.80%
26) 1,2-Dichloroethane-d4	5.31	65	179518	46.92	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.84%
32) Benzene-d6	5.34	84	622348	50.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.38%
36) 1,2-Dichloropropane-d6	6.33	67	190996	51.71	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.42%
41) Toluene-d8	7.56	98	583331	46.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.12%
43) trans-1,3-Dichloropropene-	7.85	79	85451	43.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.24%
47) 2-Hexanone-d5	8.31	63	191037	92.88	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.88%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	283989	45.50	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.00%
64) 1,2-Dichlorobenzene-d4	11.86	152	236850	46.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.72%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU030419\
Data File : VU029804.D
Acq On : 04 Mar 2019 23:28
Operator : JC/SP
Sample : K1661-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Mar 05 03:01:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM021519WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 05 01:28:42 2019
Response via : Initial Calibration

