

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030420\
 Data File : VU037035.D
 Acq On : 04 Mar 2020 11:20
 Operator : JC/MD
 Sample : L1729-21
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Mar 05 01:07:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 05 00:57:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	248843	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	236888	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	84431	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	92149	44.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.20%
7) Chloroethane-d5	1.93	69	89040	51.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	103.00%
11) 1,1-Dichloroethene-d2	2.59	63	122477	35.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.00%
21) 2-Butanone-d5	4.67	46	154148	109.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	109.97%
24) Chloroform-d	5.11	84	159981	50.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.26%
26) 1,2-Dichloroethane-d4	5.74	65	114967	53.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.48%
32) Benzene-d6	5.76	84	350193	51.12	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.24%
36) 1,2-Dichloropropane-d6	6.72	67	117244	52.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.66%
41) Toluene-d8	7.93	98	307340	48.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.72%
43) trans-1,3-Dichloropropene-	8.20	79	48931	47.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.88%
47) 2-Hexanone-d5	8.66	63	104241	96.06	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.06%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	161985	51.77	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.54%
64) 1,2-Dichlorobenzene-d4	12.21	152	91063	55.17	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.34%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030420\
Data File : VU037035.D
Acq On : 04 Mar 2020 11:20
Operator : JC/MD
Sample : L1729-21
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Mar 05 01:07:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 05 00:57:10 2020
Response via : Initial Calibration

