

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030520\
 Data File : VU037085.D
 Acq On : 05 Mar 2020 15:50
 Operator : JC/MD
 Sample : L1780-15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0DJ2

Quant Time: Mar 06 00:57:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 05 22:35:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	250200	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	238017	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	84985	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	78303	37.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.54%
7) Chloroethane-d5	1.93	69	80376	46.23	ug/L	0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.46%
11) 1,1-Dichloroethene-d2	2.59	63	111578	31.71	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.42%
21) 2-Butanone-d5	4.67	46	148200	105.15	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.15%
24) Chloroform-d	5.11	84	156703	48.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.68%
26) 1,2-Dichloroethane-d4	5.74	65	109969	50.65	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.30%
32) Benzene-d6	5.76	84	329256	47.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.68%
36) 1,2-Dichloropropane-d6	6.72	67	111016	49.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.56%
41) Toluene-d8	7.92	98	290608	45.51	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.02%
43) trans-1,3-Dichloropropene-	8.20	79	44233	42.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.36%
47) 2-Hexanone-d5	8.66	63	98835	90.65	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.65%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	157597	50.13	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	100.26%
64) 1,2-Dichlorobenzene-d4	12.21	152	88508	53.28	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.56%

Target Compounds					Ovalue
13) Acetone	2.66	43	3524	2.431 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030520\
Data File : VU037085.D
Acq On : 05 Mar 2020 15:50
Operator : JC/MD
Sample : L1780-15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0DJ2

Quant Time: Mar 06 00:57:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 05 22:35:59 2020
Response via : Initial Calibration

