


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030521\
Data File : VU042498.D
Acq On : 05 Mar 2021 11:25
Operator : SY/MD
Sample : M1570-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F1G14

Manual Integrations
APPROVED

SAM
3/8/2021 6:38:54 PM

Quant Time: Mar 08 18:14:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 05 00:31:48 2021
Response via : Initial Calibration



Quantitation Report (Qedit)


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030521\
Data File : VU042498.D
Acq On : 05 Mar 2021 11:25
Operator : SY/MD
Sample : M1570-14
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F1G14

Manual Integrations
APPROVED

SAM
3/8/2021 6:38:54 PM

Quant Time: Mar 08 18:14:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 05 00:31:48 2021
Response via : Initial Calibration



(13) Acetone (T)

2.658min (-0.000) 2.98 ug/L

response 4832

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	27.90	28.58
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030521\
 Data File : VU042498.D
 Acq On : 05 Mar 2021 11:25
 Operator : SY/MD
 Sample : M1570-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F1G14

Manual Integrations
 APPROVED

SAM
 3/8/2021 6:38:54 PM

Quant Time: Mar 08 18:14:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 05 00:31:48 2021
 Response via : Initial Calibration



(13) Acetone (T)

2.658min (-0.000) 4.42 ug/L m

response 7163

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	27.90	19.28
0.00	0.00	0.00
0.00	0.00	0.00

MD
 3/10/21

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030521\
 Data File : VU042498.D
 Acq On : 05 Mar 2021 11:25
 Operator : SY/MD
 Sample : M1570-14
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampled :
 F1G14

Manual Integrations
 APPROVED

SAM
 3/8/2021 6:38:54 PM

Quant Time: Mar 08 18:14:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 05 00:31:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	96136	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	93277	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	45854	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	23760	3.71	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	74.20%	
7) Chloroethane-d5	1.919	69	21120	4.35	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	87.00%	
11) 1,1-Dichloroethene-d2	2.575	63	41883	2.89	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	57.80%#	
20) 2-Butanone-d5	4.642	46	84036	39.94	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	79.88%	
24) Chloroform-d	5.073	84	54764	4.32	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	86.40%	
26) 1,2-Dichloroethane-d4	5.713	65	35235	4.52	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.40%	
32) Benzene-d6	5.735	84	96773	4.23	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.60%	
36) 1,2-Dichloropropane-d6	6.700	67	32642	4.42	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	88.40%	
41) Toluene-d8	7.906	98	89295	4.00	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	80.00%	
43) trans-1,3-Dichloroprop...	8.189	79	13558	3.93	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	78.60%	
46) 2-Hexanone-d5	8.645	63	70683	42.83	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	85.66%	
57) 1,1,2,2-Tetrachloroeth...	10.764	84	28768	4.38	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	87.60%	
64) 1,2-Dichlorobenzene-d4	12.201	152	37935	4.56	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	91.20%	
Target Compounds						
13) Acetone	2.658	43	7163m	4.42	ug/L	Qvalue
42) Toluene	7.977	91	1257	0.04	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed