

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030524\
 Data File : VU057982.D
 Acq On : 05 Mar 2024 10:02
 Operator : MD/SY
 Sample : VSTD0.0501
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.05001

Quant Time: Mar 06 01:28:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 01:28:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.257	114	5435	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	8087	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2587	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	326	0.051	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.709	65	199	0.053	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.687	67	202	0.046	ug/L	0.00
8) Toluene-d8	7.903	98	619	0.051	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.745	84	132	0.047	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	313	0.051	ug/L #	75
6) Trichloroethene	6.544	95	219	0.052	ug/L	95
9) 1,2-Dibromoethane	8.923	107	99	0.046	ug/L #	81
12) 1,2,3-Trichloropropane	10.814	75	100	0.048	ug/L	92
13) 1,2-Dibromo-3-chloropr...	12.990	75	24	0.057	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030524\
Data File : VU057982.D
Acq On : 05 Mar 2024 10:02
Operator : MD/SY
Sample : VSTD0.0501
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.05001

Quant Time: Mar 06 01:28:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 01:28:28 2024
Response via : Initial Calibration

