

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030524\
 Data File : VU057986.D
 Acq On : 05 Mar 2024 12:06
 Operator : MD/SY
 Sample : VSTD00205
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD002005

Quant Time: Mar 06 01:29:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 01:28:28 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	5144	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	7664	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2412	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	12721	2.117	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.685	65	7553	2.121	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	9371	2.241	ug/L	0.00
8) Toluene-d8	7.887	98	24551	2.144	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	5922	2.245	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	12494	2.131	ug/L	98
6) Trichloroethene	6.544	95	8392	2.098	ug/L	98
9) 1,2-Dibromoethane	8.924	107	4621	2.258	ug/L #	96
12) 1,2,3-Trichloropropane	10.814	75	4315	2.231	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.990	75	844	2.136	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030524\
Data File : VU057986.D
Acq On : 05 Mar 2024 12:06
Operator : MD/SY
Sample : VSTD00205
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD002005

Quant Time: Mar 06 01:29:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 01:28:28 2024
Response via : Initial Calibration

