

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030524\
 Data File : VU057990.D
 Acq On : 05 Mar 2024 15:34
 Operator : MD/SY
 Sample : VSTDCCC0.5EC
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5081

Quant Time: Mar 06 01:48:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM030524.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 01:33:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	5148	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	7598	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	2396	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	3197	0.532	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 106.000%			
4) 1,2-Dichloroethane-d4	5.685	65	1913	0.537	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 108.000%			
7) 1,2-Dichloropropane-d6	6.687	67	2336	0.563	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 112.000%			
8) Toluene-d8	7.887	98	5986	0.527	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 106.000%			
10) 1,1,2,2-Tetrachloroeth...	10.745	84	1458	0.558	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 112.000%			
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.597	62	3138	0.535	ug/L	100
6) Trichloroethene	6.544	95	2082	0.525	ug/L	99
9) 1,2-Dibromoethane	8.924	107	1131	0.557	ug/L	99
12) 1,2,3-Trichloropropane	10.814	75	1050	0.547	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.990	75	190	0.484	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030524\
Data File : VU057990.D
Acq On : 05 Mar 2024 15:34
Operator : MD/SY
Sample : VSTDCCC0.5EC
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5081

Quant Time: Mar 06 01:48:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM030524.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 01:33:52 2024
Response via : Initial Calibration

