

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030619\
 Data File : VU029826.D
 Acq On : 05 Mar 2019 17:02
 Operator : JC/SP
 Sample : K1660-02DL 5X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CWJY2DL

Quant Time: Mar 06 03:10:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Mar 06 02:51:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	238490	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	228086	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	113050	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	82053	5.27	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	105.40%
7) Chloroethane-d5	1.68	69	68813	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	2.27	63	112713	3.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.40%
20) 2-Butanone-d5	4.20	46	183876	52.22	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	104.44%
24) Chloroform-d	4.65	84	139294	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
26) 1,2-Dichloroethane-d4	5.31	65	69668	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
32) Benzene-d6	5.34	84	299449	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	6.33	67	90944	5.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.80%
41) Toluene-d8	7.56	98	272270	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
43) trans-1,3-Dichloropropene-	7.85	79	33479	4.79	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.80%
46) 2-Hexanone-d5	8.31	63	175287	51.13	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.26%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	72516	4.74	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.80%
64) 1,2-Dichlorobenzene-d4	11.86	152	109763	4.94	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.80%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	4.23	96	65240	3.700 ug/L	95
34) Trichloroethene	6.18	95	245016	14.561 ug/L	98
47) Tetrachloroethene	8.22	164	5941	0.380 ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030619\
Data File : VU029826.D
Acq On : 05 Mar 2019 17:02
Operator : JC/SP
Sample : K1660-02DL 5X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CWJY2DL

Quant Time: Mar 06 03:10:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR022119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Mar 06 02:51:24 2019
Response via : Initial Calibration

