

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030620\
 Data File : VU037101.D
 Acq On : 06 Mar 2020 11:56
 Operator : JC/MD
 Sample : VU0306WBL01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK41

Quant Time: Mar 07 01:07:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 06 23:54:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	262551	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	247150	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	83403	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	93463	42.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.78%
7) Chloroethane-d5	1.93	69	85660	46.95	ug/L	0.01
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.90%
11) 1,1-Dichloroethene-d2	2.59	63	123189	33.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.72%
21) 2-Butanone-d5	4.67	46	141584	95.73	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	95.73%
24) Chloroform-d	5.11	84	156005	46.34	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.68%
26) 1,2-Dichloroethane-d4	5.74	65	106314	46.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.32%
32) Benzene-d6	5.76	84	337097	47.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.32%
36) 1,2-Dichloropropane-d6	6.72	67	112319	48.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.02%
41) Toluene-d8	7.92	98	295568	44.58	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.16%
43) trans-1,3-Dichloropropene-	8.20	79	45617	42.39	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.78%
47) 2-Hexanone-d5	8.66	63	93179	82.30	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.30%
57) 1,1,2,2-Tetrachloroethane-	10.77	84	150903	46.22	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.44%
64) 1,2-Dichlorobenzene-d4	12.21	152	80261	49.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.46%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU030620\
Data File : VU037101.D
Acq On : 06 Mar 2020 11:56
Operator : JC/MD
Sample : VU0306WBL01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK41

Quant Time: Mar 07 01:07:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM022420WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 06 23:54:55 2020
Response via : Initial Calibration

