

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU030719\
Data File : VU029837.D
Acq On : 06 Mar 2019 13:40
Operator : JC/SP
Sample : VU0307WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0307WBL01

Quant Time: Mar 08 01:40:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U030719DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Mar 08 00:30:10 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.74	96	62882	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.31	95	18390	0.81	ug/l	0.00
Spiked Amount	1.000		Recovery	=	81.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	21386	0.82	ug/l	0.00
Spiked Amount	1.000		Recovery	=	82.00%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU030719\
Data File : VU029837.D
Acq On : 06 Mar 2019 13:40
Operator : JC/SP
Sample : VU0307WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0307WBL01

Quant Time: Mar 08 01:40:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\524U030719DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Mar 08 00:30:10 2019
Response via : Initial Calibration

