

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU030719\
 Data File : VU029842.D
 Acq On : 06 Mar 2019 16:30
 Operator : JC/SP
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 08 01:56:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U030719DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Mar 08 00:30:10 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 i	Fluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.335	0.330	1.5	101	0.00
3 t	Chloromethane	0.298	0.283	5.0	98	0.00
4 Rt	Vinyl Chloride	0.335	0.337	-0.6	101	0.00
5 T	Bromomethane	0.229	0.181	21.0	92	0.00
6 T	Chloroethane	0.201	0.200	0.5	101	0.00
7 T	Trichlorofluoromethane	0.457	0.457	0.0	101	0.00
8	1,1,2-Trichloro-1,2,2-trifl	0.266	0.261	1.9	99	0.00
9 Rt	1,1-Dichloroethene	0.260	0.262	-0.8	101	0.00
10 t	Iodomethane	0.226	0.299	-32.3#	101	0.00
11 t	Allyl Chloride	0.366	0.358	2.2	100	0.00
12 t	Acrylonitrile	0.057	0.055	3.5	98	0.00
13 T	Acetone	0.047	0.045	4.3	105	0.00
14 T	Carbon Disulfide	0.842	0.827	1.8	99	0.00
15 RT	Methylene Chloride	0.312	0.297	4.8	102	0.00
16 RT	trans-1,2-Dichloroethene	0.294	0.289	1.7	100	0.00
17 t	1,1-Dichloroethane	0.492	0.506	-2.8	106	0.00
18 T	2-Butanone	0.064	0.067	-4.7	100	0.00
19	Cyclohexane	0.382	0.398	-4.2	100	0.00
20	Methylcyclohexane	0.445	0.472	-6.1	100	0.00
21 T	2,2-Dichloropropane	0.402	0.399	0.7	101	0.00
22 RT	cis-1,2-Dichloroethene	0.311	0.318	-2.3	101	0.00
23 t	Diethyl Ether	0.195	0.196	-0.5	100	0.00
24 t	tert-Butyl Alcohol	0.020	0.020	0.0	101	0.00
25 t	Methyl tert-Butyl Ether	0.655	0.669	-2.1	103	0.00
26 t	Bromochloromethane	0.138	0.142	-2.9	104	0.00
27 t	Chloroform	0.498	0.500	-0.4	102	0.00
28 RT	1,1,1-Trichloroethane	0.422	0.429	-1.7	101	0.00
29 T	1,1-Dichloropropene	0.375	0.381	-1.6	100	0.00
30 RT	Carbon Tetrachloride	0.368	0.381	-3.5	101	0.00
31 t	Isopropyl Ether	0.678	0.698	-2.9	103	0.00
32	Ethyl-t-butyl ether	0.659	0.686	-4.1	102	0.00
33	Tert-Amyl methyl ether	0.610	0.653	-7.0	103	0.00
34 t	Propionitrile	0.019	0.021	-10.5	99	0.00
35 RT	Benzene	1.105	1.136	-2.8	103	0.00
36 RT	1,2-Dichloroethane	0.302	0.303	-0.3	100	0.00
37 RT	Trichloroethene	0.323	0.324	-0.3	101	0.00
38 Rt	1,2-Dichloropropane	0.280	0.289	-3.2	102	0.00
39 t	Methacrylonitrile	0.080	0.083	-3.8	101	0.00
40 t	Methyl acrylate	0.126	0.131	-4.0	100	0.00
41 t	Tetrahydrofuran	0.042	0.042	0.0	102	0.00
42 t	1-Chlorobutane	0.479	0.501	-4.6	102	0.00
43 T	Dibromomethane	0.145	0.150	-3.4	101	0.00
44 T	Bromodichloromethane	0.356	0.369	-3.7	102	0.00
45 T	4-Methyl-2-Pentanone	0.151	0.163	-7.9	103	0.00
46 t	t-1,4-Dichloro-2-butene	0.063	0.071	-12.7	101	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU030719\
 Data File : VU029842.D
 Acq On : 06 Mar 2019 16:30
 Operator : JC/SP
 Sample : VSTDCCC010
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC010

Quant Time: Mar 08 01:56:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U030719DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Fri Mar 08 00:30:10 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 200%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
47 t	Methyl methacrylate	0.124	0.138	-11.3	103	0.00
48 t	Ethyl methacrylate	0.238	0.278	-16.8	104	0.00
49 Rt	Toluene	0.679	0.728	-7.2	101	0.00
50 T	t-1,3-Dichloropropene	0.312	0.337	-8.0	101	0.00
51 T	cis-1,3-Dichloropropene	0.386	0.414	-7.3	102	0.00
52 RT	1,1,2-Trichloroethane	0.215	0.220	-2.3	104	0.00
53 t	1,3-Dichloropropane	0.352	0.365	-3.7	103	0.00
54 t	2-Hexanone	0.105	0.115	-9.5	101	0.00
55 t	Dibromochloromethane	0.255	0.271	-6.3	102	0.00
56 T	1,2-Dibromoethane	0.203	0.207	-2.0	101	0.00
57 S	4-Bromofluorobenzene	0.360	0.385	-6.9	102	0.00
58 RT	Tetrachloroethene	0.330	0.333	-0.9	102	0.00
59 Rt	Chlorobenzene	0.780	0.821	-5.3	101	0.00
60 T	1,1,1,2-Tetrachloroethane	0.280	0.289	-3.2	102	0.00
61 t	Pentachloroethane	0.195	0.204	-4.6	100	0.00
62 t	Hexachloroethane	0.205	0.219	-6.8	98	0.00
63 Rt	Ethyl Benzene	1.278	1.394	-9.1	102	0.00
64 RT	m/p-Xylenes	0.505	0.563	-11.5	102	0.00
65 RT	o-Xylene	0.485	0.543	-12.0	103	0.00
66 RT	Styrene	0.806	0.917	-13.8	103	0.00
67 t	Bromoform	0.150	0.164	-9.3	102	0.00
68 S	1,2-Dichlorobenzene-d4	0.417	0.434	-4.1	102	0.00
69 T	Isopropylbenzene	1.286	1.430	-11.2	103	0.00
70 T	1,1,2,2-Tetrachloroethane	0.262	0.278	-6.1	105	0.00
71 T	1,2,3-Trichloropropane	0.194	0.198	-2.1	102	0.00
72 t	Bromobenzene	0.340	0.367	-7.9	103	0.00
73 t	n-propylbenzene	0.362	0.407	-12.4	102	0.00
74 t	2-Chlorotoluene	0.321	0.348	-8.4	102	0.00
75 t	1,3,5-Trimethylbenzene	1.084	1.215	-12.1	101	0.00
76 t	4-Chlorotoluene	0.327	0.359	-9.8	101	0.00
77 t	tert-Butylbenzene	1.072	1.189	-10.9	101	0.00
78 t	1,2,4-Trimethylbenzene	1.095	1.243	-13.5	102	0.00
79 t	sec-Butylbenzene	1.421	1.613	-13.5	103	0.00
80	Nitrobenzene	0.005	0.005	0.0	94	0.00
81 t	p-Isopropyltoluene	1.189	1.360	-14.4	102	0.00
82 t	1,3-Dichlorobenzene	0.674	0.721	-7.0	102	0.00
83 Rt	1,4-Dichlorobenzene	0.659	0.714	-8.3	103	0.00
84 t	n-Butylbenzene	1.084	1.244	-14.8	103	0.00
85 Rt	1,2-Dichlorobenzene	0.636	0.671	-5.5	102	0.00
86 t	1,2-Dibromo-3-Chloropropane	0.037	0.041	-10.8	100	0.00
87 Rt	1,2,4-Trichlorobenzene	0.407	0.462	-13.5	102	0.00
88 t	Hexachlorobutadiene	0.256	0.271	-5.9	102	0.00
89 t	Naphthalene	0.651	0.784	-20.4	100	0.00
90 t	1,2,3-Trichlorobenzene	0.383	0.427	-11.5	100	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU030719\
Data File : VU029842.D
Acq On : 06 Mar 2019 16:30
Operator : JC/SP
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC010

Quant Time: Mar 08 01:56:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\524U030719DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Mar 08 00:30:10 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			