

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU030819\
Data File : VU029845.D
Acq On : 07 Mar 2019 10:20
Operator : JC/SP
Sample : VU0308WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0308WBL01

Quant Time: Mar 08 02:17:34 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U030719DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Mar 08 00:30:10 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	5.74	96	60833	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.31	95	19271	0.88	ug/l	0.00
Spiked Amount	1.000		Recovery	=	88.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	21264	0.84	ug/l	0.00
Spiked Amount	1.000		Recovery	=	84.00%	


Target Compounds	Ovalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU030819\
Data File : VU029845.D
Acq On : 07 Mar 2019 10:20
Operator : JC/SP
Sample : VU0308WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0308WBL01

Quant Time: Mar 08 02:17:34 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U030719DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Mar 08 00:30:10 2019
Response via : Initial Calibration





#1

Fluorobenzene

Concen: 1.000 ug/l

RT: 5.74 min Scan# 1446

Delta R.T. -0.00 min

Lab File: VU029845.D

Acq: 07 Mar 2019 10:20

Instrument :

MSVOA_U

ClientSampleId :

VU0308WBL01

Tot Ion: 96 Resp: 60833


Ion Ratio Lower Upper

96 100

70 18.2 14.4 21.6

95 9.0 7.0 10.6

97 6.0 0.0 0.0#






Abundance Ion 96.00 (95.70 to 96.70): VU0

Ion 70.00 (69.70 to 70.70): VU0

Ion 95.00 (94.70 to 95.70): VU0

Ion 97.00 (96.70 to 97.70): VU0



#57

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 10.31 min Scan# 2867

Delta R.T. 0.00 min

Lab File: VU029845.D

Acq: 07 Mar 2019 10:20


Tgt Ion: 95 Resp: 19271

Ion Ratio Lower Upper

95 100

174 89.4 76.5 114.7



176 87.8 74.2 111.2




Abundance Ion 95.00 (94.70 to 95.70): VU0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V






#68
 1,2-Dichlorobenzene-d4
 Concen: N.D. ug/l
 RT: 11.86 min Scan# 3349
 Delta R.T. 0.00 min
 Lab File: VU029845.D
 Acq: 07 Mar 2019 10:20

Instrument :
 MSVOA_U
 ClientSampleId :
 VU0308WBL01

Tot Ion:152 Resp: 21264
 Ion Ratio Lower Upper
 152 100
 115 59.2 0.0 124.0
 150 157.8 0.0 631.2



Abundance Ion 152.00 (151.70 to 152.70): V
 Ion 115.00 (114.70 to 115.70): V
 Ion 150.00 (149.70 to 150.70): V

