

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030821\
 Data File : VU042525.D
 Acq On : 08 Mar 2021 13:41
 Operator : SY/MD
 Sample : M1588-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampled :
 C03T5

Quant Time: Mar 09 01:37:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 09 01:35:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.256	114	96341	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	91347	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	44310	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	24044	3.74	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery =	74.80%		
7) Chloroethane-d5	1.919	69	21406	4.40	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery =	88.00%		
11) 1,1-Dichloroethene-d2	2.575	63	52214	3.60	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery =	72.00%		
20) 2-Butanone-d5	4.646	46	103228	48.95	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery =	97.90%		
24) Chloroform-d	5.073	84	55410	4.36	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery =	87.20%		
26) 1,2-Dichloroethane-d4	5.713	65	37658	4.82	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery =	96.40%		
32) Benzene-d6	5.736	84	101420	4.53	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery =	90.60%		
36) 1,2-Dichloropropane-d6	6.697	67	33951	4.70	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery =	94.00%		
41) Toluene-d8	7.906	98	95766	4.38	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery =	87.60%		
43) trans-1,3-Dichloroprop...	8.186	79	14315	4.24	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery =	84.80%		
46) 2-Hexanone-d5	8.645	63	72652	44.95	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery =	89.90%		
57) 1,1,2,2-Tetrachloroeth...	10.764	84	29495	4.58	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery =	91.60%		
64) 1,2-Dichlorobenzene-d4	12.202	152	38631	4.81	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery =	96.20%		

Target Compounds						Qvalue
3) Chloromethane	1.523	50	1192	0.14	ug/L	93
12) 1,1-Dichloroethene	2.588	96	17205	2.53	ug/L #	71
13) Acetone	2.655	43	1384	0.85	ug/L	70
19) 1,1-Dichloroethane	3.880	63	8537	0.62	ug/L	94
29) 1,1,1-Trichloroethane	5.327	97	86029	6.77	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030821\
 Data File : VU042525.D
 Acq On : 08 Mar 2021 13:41
 Operator : SY/MD
 Sample : M1588-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleID :
 C03T5

Quant Time: Mar 09 01:37:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Mar 09 01:35:43 2021
 Response via : Initial Calibration

