

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030824\
 Data File : VU058006.D
 Acq On : 08 Mar 2024 10:45
 Operator : MD/SY
 Sample : P1681-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COU30

Quant Time: Mar 09 00:11:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 09 00:10:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	166365	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	155083	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	70675	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	49414	4.128	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	82.600%
7) Chloroethane-d5	1.907	69	42677	4.338	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.800%
11) 1,1-Dichloroethene-d2	2.560	65	22124	4.271	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.400%
20) 2-Butanone-d5	4.612	46	98962	50.178	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.360%
24) Chloroform-d	5.052	84	101559	4.486	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
26) 1,2-Dichloroethane-d4	5.695	65	49257	4.802	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
32) Benzene-d6	5.721	84	203947	4.513	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
36) 1,2-Dichloropropane-d6	6.685	67	63495	4.621	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.400%
41) Toluene-d8	7.894	98	178649	4.363	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
43) trans-1,3-Dichloroprop...	8.177	79	20019	4.237	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.627	63	79558	50.154	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.300%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	39973	4.745	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%
66) 1,2-Dichlorobenzene-d4	12.190	152	55373	4.757	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.200%
Target Compounds					Qvalue	
13) Acetone	2.615	43	32371	22.055	ug/L	99
21) 2-Butanone	4.711	43	12514	5.411	ug/L	77
42) Toluene	7.968	91	18895	0.381	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030824\
Data File : VU058006.D
Acq On : 08 Mar 2024 10:45
Operator : MD/SY
Sample : P1681-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0U30

Quant Time: Mar 09 00:11:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 09 00:10:03 2024
Response via : Initial Calibration

