

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU030824\  
 Data File : VU058023.D  
 Acq On : 08 Mar 2024 17:32  
 Operator : MD/SY  
 Sample : P1636-17  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 DCNS4

Quant Time: Mar 09 00:15:05 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR022924WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Sat Mar 09 00:10:03 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon           | Response | Conc   | Units    | Dev(Min) |
|-------------------------------|--------|----------------|----------|--------|----------|----------|
| -----                         |        |                |          |        |          |          |
| Internal Standards            |        |                |          |        |          |          |
| 1) 1,4-Difluorobenzene        | 6.242  | 114            | 138882   | 5.000  | ug/L     | 0.00     |
| 28) Chlorobenzene-d5          | 9.412  | 117            | 134155   | 5.000  | ug/L     | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.807 | 152            | 60419    | 5.000  | ug/L     | 0.00     |
|                               |        |                |          |        |          |          |
| System Monitoring Compounds   |        |                |          |        |          |          |
| 4) Vinyl Chloride-d3          | 1.592  | 65             | 42130    | 4.216  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 40 - 130 | Recovery | =      | 84.400%  |          |
| 7) Chloroethane-d5            | 1.907  | 69             | 38295    | 4.663  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 130 | Recovery | =      | 93.200%  |          |
| 11) 1,1-Dichloroethene-d2     | 2.560  | 65             | 19282    | 4.459  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 125 | Recovery | =      | 89.200%  |          |
| 20) 2-Butanone-d5             | 4.612  | 46             | 93131    | 56.566 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 40 - 130 | Recovery | =      | 113.140% |          |
| 24) Chloroform-d              | 5.052  | 84             | 92061    | 4.871  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 97.400%  |          |
| 26) 1,2-Dichloroethane-d4     | 5.692  | 65             | 43793    | 5.114  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 102.200% |          |
| 32) Benzene-d6                | 5.718  | 84             | 184604   | 4.722  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 125 | Recovery | =      | 94.400%  |          |
| 36) 1,2-Dichloropropane-d6    | 6.682  | 67             | 59030    | 4.967  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 60 - 140 | Recovery | =      | 99.400%  |          |
| 41) Toluene-d8                | 7.891  | 98             | 156592   | 4.421  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 70 - 130 | Recovery | =      | 88.400%  |          |
| 43) trans-1,3-Dichloroprop... | 8.177  | 79             | 19379    | 4.742  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 55 - 130 | Recovery | =      | 94.800%  |          |
| 46) 2-Hexanone-d5             | 8.627  | 63             | 77410    | 56.413 | ug/L     | 0.00     |
| Spiked Amount                 | 50.000 | Range 45 - 130 | Recovery | =      | 112.820% |          |
| 56) 1,1,2,2-Tetrachloroeth... | 10.750 | 84             | 36901    | 5.063  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 65 - 120 | Recovery | =      | 101.200% |          |
| 66) 1,2-Dichlorobenzene-d4    | 12.190 | 152            | 50649    | 5.090  | ug/L     | 0.00     |
| Spiked Amount                 | 5.000  | Range 80 - 120 | Recovery | =      | 101.800% |          |
|                               |        |                |          |        |          |          |
| Target Compounds              |        |                |          |        |          | Qvalue   |
| 22) cis-1,2-Dichloroethene    | 4.657  | 96             | 8336     | 0.830  | ug/L     | 97       |
| 47) Tetrachloroethene         | 8.547  | 164            | 66652    | 8.869  | ug/L     | 97       |
| -----                         |        |                |          |        |          |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU030824\  
Data File : VU058023.D  
Acq On : 08 Mar 2024 17:32  
Operator : MD/SY  
Sample : P1636-17  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
DCNS4

Quant Time: Mar 09 00:15:05 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\SFAMUTR022924WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Sat Mar 09 00:10:03 2024  
Response via : Initial Calibration

