

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030824\
 Data File : VU058018.D
 Acq On : 08 Mar 2024 15:33
 Operator : MD/SY
 Sample : P1636-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DCNQ4

Quant Time: Mar 09 00:13:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Mar 09 00:10:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	139197	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	135400	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	61681	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	43334	4.327	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.600%
7) Chloroethane-d5	1.904	69	39193	4.761	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.557	65	19714	4.549	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.000%
20) 2-Butanone-d5	4.612	46	92441	56.020	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.040%
24) Chloroform-d	5.052	84	93373	4.929	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
26) 1,2-Dichloroethane-d4	5.695	65	45348	5.284	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.600%
32) Benzene-d6	5.718	84	183565	4.652	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
36) 1,2-Dichloropropane-d6	6.685	67	59848	4.989	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.800%
41) Toluene-d8	7.891	98	157583	4.408	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
43) trans-1,3-Dichloroprop...	8.174	79	18970	4.599	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.000%
46) 2-Hexanone-d5	8.627	63	78738	56.853	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.700%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	39036	5.307	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	50961	5.017	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.400%
Target Compounds						Qvalue
17) Methyl tert-butyl Ether	3.354	73	7766	0.391	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030824\
Data File : VU058018.D
Acq On : 08 Mar 2024 15:33
Operator : MD/SY
Sample : P1636-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DCNQ4

Quant Time: Mar 09 00:13:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Mar 09 00:10:03 2024
Response via : Initial Calibration





#17
Methyl tert-butyl Ether
Concen: 0.391 ug/L
RT: 3.354 min Scan# 64
Delta R.T. 0.003 min
Lab File: VU058018.D
Acq: 08 Mar 2024 15:33

Instrument :
MSVOA_U
ClientSampleId :
DCNQ4

Tgt Ion:	73	Resp:	7766
Ion	Ratio	Lower	Upper
73	100		
43	18.5	16.2	24.4
57	20.6	17.9	26.9

