

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030920\
 Data File : VU037138.D
 Acq On : 09 Mar 2020 18:03
 Operator : JC/MD
 Sample : L1161-07 0.4PPB
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 LOD-MDL-WATER-01-QT1-2020

Quant Time: Mar 10 08:57:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U030920DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Mon Mar 09 18:12:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	6.15	96	15221	1.00	ug/l	0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.65	95	4929	0.84	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.00%	
68) 1,2-Dichlorobenzene-d4	12.21	152	5152	0.86	ug/l	0.00
Spiked Amount 1.000			Recovery	=	86.00%	
Target Compounds						
					Ovalue	
34) Propionitrile	4.83	54	601	1.616	ug/l #	67
86) 1,2-Dibromo-3-Chloropropan	13.03	75	249	0.311	ug/l #	16
-----	-----	-----	-----	-----	-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU030920\
Data File : VU037138.D
Acq On : 09 Mar 2020 18:03
Operator : JC/MD
Sample : L1161-07 0.4PPB
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
LOD-MDL-WATER-01-QT1-2020

Quant Time: Mar 10 08:57:04 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U030920DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Mar 09 18:12:43 2020
Response via : Initial Calibration

