

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030921\
 Data File : VU042541.D
 Acq On : 09 Mar 2021 16:05
 Operator : SY/MD
 Sample : MDL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MDL01

Quant Time: Mar 11 13:09:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U030921DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Mar 11 12:50:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Fluorobenzene	6.124	96	19406	1.00	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	7867	0.92	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.00%	
68) 1,2-Dichlorobenzene-d4	12.201	152	7790	0.88	ug/l	0.00
Spiked Amount 1.000			Recovery	=	88.00%	
Target Compounds						
15) Methylene Chloride	3.054	84	3434	0.59	ug/l	Qvalue 93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU030921\
Data File : VU042541.D
Acq On : 09 Mar 2021 16:05
Operator : SY/MD
Sample : MDL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MDL01

Quant Time: Mar 11 13:09:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U030921DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Mar 11 12:50:28 2021
Response via : Initial Calibration









#1
Fluorobenzene
Concen: 1.00 ug/l
RT: 6.124 min Scan# 1488
Delta R.T. 0.003 min
Lab File: VU042541.D
Acq: 09 Mar 2021 16:05

Instrument :
MSVOA_U
ClientSampleId :
MDL01

Tgt Ion: 96 Resp: 19406




Ion	Ratio	Lower	Upper
96	100		
70	21.4	16.3	24.5
95	8.6	7.0	10.4
97	6.3	0.0	0.0




#15
Methylene Chloride
Concen: 0.59 ug/l
RT: 3.054 min Scan# 533
Delta R.T. -0.000 min
Lab File: VU042541.D
Acq: 09 Mar 2021 16:05

Tgt Ion: 84 Resp: 3434

Ion	Ratio	Lower	Upper
84	100		
49	132.5	114.6	172.0
51	41.1	0.0	83.0
86	59.4	52.3	78.5









#57
4-Bromofluorobenzene
Concen: 0.92 ug/l
RT: 10.642 min Scan# 2893
Delta R.T. 0.003 min
Lab File: VU042541.D
Acq: 09 Mar 2021 16:05

Instrument :
MSVOA_U
ClientSampleId :
MDL01

Tgt Ion: 95 Resp: 7867
Ion Ratio Lower Upper
95 100
174 95.0 73.1 109.7
176 85.2 63.4 95.2



#68
1,2-Dichlorobenzene-d4
Concen: 0.88 ug/l
RT: 12.201 min Scan# 3378
Delta R.T. -0.000 min
Lab File: VU042541.D
Acq: 09 Mar 2021 16:05

Tgt Ion: 152 Resp: 7790
Ion Ratio Lower Upper
152 100
115 67.5 0.0 266.0
150 166.4 0.0 657.2

