

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031019\
Data File : VU029876.D
Acq On : 08 Mar 2019 09:39
Operator : JC/SP
Sample : VU0310WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0310WBL01

Quant Time: Mar 08 22:48:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U030719DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Mar 08 00:30:10 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	62527	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	19048	0.85	ug/l	0.00
Spiked Amount 1.000			Recovery	=	85.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	21409	0.82	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.00%	

Target Compounds	Ovalue
-----	-----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031019\
Data File : VU029876.D
Acq On : 08 Mar 2019 09:39
Operator : JC/SP
Sample : VU0310WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0310WBL01

Quant Time: Mar 08 22:48:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U030719DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Fri Mar 08 00:30:10 2019
Response via : Initial Calibration

