

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031020\
Data File : VU037144.D
Acq On : 10 Mar 2020 10:42
Operator : JC/MD
Sample : VU0310WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0310WBL01

Quant Time: Mar 10 11:42:51 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U030920DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Mar 09 15:46:40 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Fluorobenzene	6.15	96	13682	1.00	ug/l	0.00

System Monitoring Compounds

57) 4-Bromofluorobenzene	10.65	95	4373	0.83	ug/l	0.00
Spiked Amount	1.000		Recovery	=	83.00%	
68) 1,2-Dichlorobenzene-d4	12.21	152	4475	0.83	ug/l	0.00
Spiked Amount	1.000		Recovery	=	83.00%	

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031020\
Data File : VU037144.D
Acq On : 10 Mar 2020 10:42
Operator : JC/MD
Sample : VU0310WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0310WBL01

Quant Time: Mar 10 11:42:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U030920DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Mon Mar 09 15:46:40 2020
Response via : Initial Calibration

