

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031021\
 Data File : VU042549.D
 Acq On : 10 Mar 2021 11:17
 Operator : SY/MD
 Sample : VU0310WBL04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK26

Quant Time: Mar 11 13:47:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U030921DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Mar 11 12:50:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.124	96	18998	1.00	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	7955	0.95	ug/l	0.00
Spiked Amount 1.000			Recovery	=	95.00%	
68) 1,2-Dichlorobenzene-d4	12.205	152	7186	0.83	ug/l	0.00
Spiked Amount 1.000			Recovery	=	83.00%	
Target Compounds						
15) Methylene Chloride	3.060	84	1031	0.18	ug/l	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031021\
Data File : VU042549.D
Acq On : 10 Mar 2021 11:17
Operator : SY/MD
Sample : VU0310WBL04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK26

Quant Time: Mar 11 13:47:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U030921DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Mar 11 12:50:28 2021
Response via : Initial Calibration

