

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU031021\  
 Data File : VU042550.D  
 Acq On : 10 Mar 2021 11:47  
 Operator : SY/MD  
 Sample : VU0310WBL05  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VBLK27

Quant Time: Mar 11 13:54:16 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U030921DW.M  
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
 QLast Update : Thu Mar 11 12:50:28 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|------|--------|----------|
| -----                       |        |      |          |      |        |          |
| Internal Standards          |        |      |          |      |        |          |
| 1) Fluorobenzene            | 6.124  | 96   | 18791    | 1.00 | ug/l   | # 0.00   |
|                             |        |      |          |      |        |          |
| System Monitoring Compounds |        |      |          |      |        |          |
| 57) 4-Bromofluorobenzene    | 10.642 | 95   | 6695     | 0.81 | ug/l   | 0.00     |
| Spiked Amount 1.000         |        |      | Recovery | =    | 81.00% |          |
| 68) 1,2-Dichlorobenzene-d4  | 12.201 | 152  | 6873     | 0.80 | ug/l   | 0.00     |
| Spiked Amount 1.000         |        |      | Recovery | =    | 80.00% |          |
|                             |        |      |          |      |        |          |
| Target Compounds            |        |      |          |      |        | Qvalue   |
| 15) Methylene Chloride      | 3.060  | 84   | 1014     | 0.18 | ug/l   | # 81     |
| -----                       |        |      |          |      |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU031021\  
Data File : VU042550.D  
Acq On : 10 Mar 2021 11:47  
Operator : SY/MD  
Sample : VU0310WBL05  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VBLK27

Quant Time: Mar 11 13:54:16 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U030921DW.M  
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
QLast Update : Thu Mar 11 12:50:28 2021  
Response via : Initial Calibration

