

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU031021\  
 Data File : VU042551.D  
 Acq On : 10 Mar 2021 12:43  
 Operator : SY/MD  
 Sample : MDL03  
 Misc : 25.0mL/MSVOA\_U/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 MDL03

Quant Time: Mar 11 13:57:26 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U030921DW.M  
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
 QLast Update : Thu Mar 11 12:50:28 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc | Units  | Dev(Min)  |
|-----------------------------|--------|------|----------|------|--------|-----------|
| -----                       |        |      |          |      |        |           |
| Internal Standards          |        |      |          |      |        |           |
| 1) Fluorobenzene            | 6.124  | 96   | 17688    | 1.00 | ug/l   | # 0.00    |
| System Monitoring Compounds |        |      |          |      |        |           |
| 57) 4-Bromofluorobenzene    | 10.642 | 95   | 7379     | 0.95 | ug/l   | 0.00      |
| Spiked Amount 1.000         |        |      | Recovery | =    | 95.00% |           |
| 68) 1,2-Dichlorobenzene-d4  | 12.205 | 152  | 7618     | 0.95 | ug/l   | 0.00      |
| Spiked Amount 1.000         |        |      | Recovery | =    | 95.00% |           |
| Target Compounds            |        |      |          |      |        |           |
| 15) Methylene Chloride      | 3.054  | 84   | 1744     | 0.33 | ug/l   | Qvalue 98 |
| -----                       |        |      |          |      |        |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_U\Data\VU031021\  
Data File : VU042551.D  
Acq On : 10 Mar 2021 12:43  
Operator : SY/MD  
Sample : MDL03  
Misc : 25.0mL/MSVOA\_U/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
MDL03

Quant Time: Mar 11 13:57:26 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_U\Method\524U030921DW.M  
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER  
QLast Update : Thu Mar 11 12:50:28 2021  
Response via : Initial Calibration

