

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031021\
 Data File : VU042553.D
 Acq On : 10 Mar 2021 15:27
 Operator : SY/MD
 Sample : MDL05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MDL05

Quant Time: Mar 11 14:05:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U030921DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Thu Mar 11 12:50:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.124	96	17600	1.00	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	7112	0.92	ug/l	0.00
Spiked Amount 1.000			Recovery	=	92.00%	
68) 1,2-Dichlorobenzene-d4	12.201	152	6808	0.85	ug/l	0.00
Spiked Amount 1.000			Recovery	=	85.00%	
Target Compounds						
15) Methylene Chloride	3.057	84	1676	0.32	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031021\
Data File : VU042553.D
Acq On : 10 Mar 2021 15:27
Operator : SY/MD
Sample : MDL05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MDL05

Quant Time: Mar 11 14:05:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U030921DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Thu Mar 11 12:50:28 2021
Response via : Initial Calibration

