

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031119\
 Data File : VU029927.D
 Acq On : 09 Mar 2019 18:02
 Operator : JC/SP
 Sample : K1789-21
 Misc : 5.70µ/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF0Q4

Manual Integrations
 APPROVED

MMDadoda
 3/11/2019 10:54:30 AM

Quant Time: Mar 11 03:57:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Mar 11 02:20:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	567119	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	568560	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	303435	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	146233	42.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.52%
7) Chloroethane-d5	1.64	69	58929m	20.26	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	40.52%#
11) 1,1-Dichloroethene-d2	2.23	63	201542m	30.90	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.80%
21) 2-Butanone-d5	4.19	46	223013	83.19	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.19%
24) Chloroform-d	4.64	84	263175	39.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.16%
26) 1,2-Dichloroethane-d4	5.30	65	161585	40.95	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.90%
32) Benzene-d6	5.33	84	576364	40.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.36%
36) 1,2-Dichloropropane-d6	6.32	67	172599	40.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.66%
41) Toluene-d8	7.56	98	542060	39.87	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.74%#
43) trans-1,3-Dichloropropene-	7.85	79	81158	39.19	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.38%
47) 2-Hexanone-d5	8.32	63	211640	91.91	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.91%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	299741	44.36	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	88.72%
64) 1,2-Dichlorobenzene-d4	11.86	152	243591	40.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.86%

Target Compounds					Ovalue
3) Chloromethane	1.32	50	56559	9.880 ug/L	100
6) Bromomethane	1.60	94	3060m	2.603 ug/L	
46) Tetrachloroethene	8.22	164	61923	17.565 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031119\
Data File : VU029927.D
Acq On : 09 Mar 2019 18:02
Operator : JC/SP
Sample : K1789-21
Misc : 5.70µ/5mL/100µL/5.0mL/MSVOA_U/MEOH
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF0Q4

Manual Integrations
APPROVED

MMDadoda
3/11/2019 10:54:30 AM

Quant Time: Mar 11 03:57:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031119WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Mar 11 02:20:19 2019
Response via : Initial Calibration

