

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031119\
 Data File : VU029917.D
 Acq On : 09 Mar 2019 14:06
 Operator : JC/SP
 Sample : VU0311MBL01
 Misc : 5.00g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK52

Quant Time: Mar 11 02:41:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031119WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Mar 11 02:20:19 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	525724	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	502970	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	237098	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	150990	47.07	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.14%
7) Chloroethane-d5	1.68	69	129401	47.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.98%
11) 1,1-Dichloroethene-d2	2.27	63	216713	35.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.68%
21) 2-Butanone-d5	4.17	46	254829	102.54	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.54%
24) Chloroform-d	4.64	84	295410	47.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.64%
26) 1,2-Dichloroethane-d4	5.31	65	180903	49.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.92%
32) Benzene-d6	5.34	84	638887	50.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.94%
36) 1,2-Dichloropropane-d6	6.33	67	198851	52.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.06%
41) Toluene-d8	7.56	98	594806	49.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.92%
43) trans-1,3-Dichloropropene-	7.85	79	91852	50.14	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.28%
47) 2-Hexanone-d5	8.31	63	210985	103.58	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	103.58%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	293327	49.07	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.14%
64) 1,2-Dichlorobenzene-d4	11.86	152	236532	50.86	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.72%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031119\
Data File : VU029917.D
Acq On : 09 Mar 2019 14:06
Operator : JC/SP
Sample : VU0311MBL01
Misc : 5.00µ/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK52

Quant Time: Mar 11 02:41:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031119WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Mar 11 02:20:19 2019
Response via : Initial Calibration

