

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031120\
 Data File : VU037172.D
 Acq On : 11 Mar 2020 17:30
 Operator : JC/MD
 Sample : L1883-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 CON94

Quant Time: Mar 12 09:52:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR031120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Mar 12 01:55:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	108491	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	108472	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	36470	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	38548	5.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.20%
7) Chloroethane-d5	1.93	69	30521	6.06	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.20%
11) 1,1-Dichloroethene-d2	2.59	63	53313	3.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.60%
20) 2-Butanone-d5	4.67	46	98198	53.89	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.78%
24) Chloroform-d	5.11	84	75617	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
26) 1,2-Dichloroethane-d4	5.74	65	45302	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
32) Benzene-d6	5.76	84	140214	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.20%
36) 1,2-Dichloropropane-d6	6.72	67	46370	5.38	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.60%
41) Toluene-d8	7.93	98	122025	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.60%
43) trans-1,3-Dichloropropene-	8.20	79	17394	5.00	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.00%
46) 2-Hexanone-d5	8.66	63	72098	49.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.50%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	34903	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
64) 1,2-Dichlorobenzene-d4	12.21	152	37096	5.89	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	117.80%

Target Compounds				Ovalue
13) Acetone	2.66	43	3620	2.424 ug/L 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031120\
Data File : VU037172.D
Acq On : 11 Mar 2020 17:30
Operator : JC/MD
Sample : L1883-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
CON94

Quant Time: Mar 12 09:52:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR031120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Mar 12 01:55:55 2020
Response via : Initial Calibration

