

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031121\
 Data File : VU042557.D
 Acq On : 11 Mar 2021 10:20
 Operator : SY/MD
 Sample : VU0311WBL06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK28

Quant Time: Mar 11 10:39:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U030921DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Wed Mar 10 06:46:59 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Fluorobenzene	6.124	96	18607	1.00	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.642	95	6832	0.83	ug/l	0.00
Spiked Amount 1.000			Recovery	=	83.00%	
68) 1,2-Dichlorobenzene-d4	12.205	152	6970	0.82	ug/l	0.00
Spiked Amount 1.000			Recovery	=	82.00%	
Target Compounds						
15) Methylene Chloride	3.057	84	874	0.16	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031121\
Data File : VU042557.D
Acq On : 11 Mar 2021 10:20
Operator : SY/MD
Sample : VU0311WBL06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK28

Quant Time: Mar 11 10:39:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U030921DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Wed Mar 10 06:46:59 2021
Response via : Initial Calibration

