


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU031219\
 Data File : VU029962.D
 Acq On : 11 Mar 2019 07:55
 Operator : JC/SP
 Sample : BFB61
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB61

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031119WMA.M
 Title : VOC Analysis
 Last Update : Mon Mar 11 12:44:43 2019



AutoFind: Scans 2865, 2866, 2867; Background Corrected with Scan 2854

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.3	11220	PASS
75	95	30	80	45.2	33221	PASS
95	95	100	100	100.0	73437	PASS
96	95	5	9	7.0	5150	PASS
173	174	0.00	2	0.9	621	PASS
174	95	50	120	90.9	66776	PASS
175	174	5	9	7.0	4674	PASS
176	174	95	101	95.6	63850	PASS
177	176	5	9	6.4	4106	PASS

Scan 2854 (10.266 min): VU029962.D
BFB61

m/z	Abundance
40.00	1902.0
43.90	370.0

Instrument :
MSVOA_U
ClientSampleId :
BFB61

m/z	Abundance
36.00	468.0
37.00	2407.0
38.00	2586.0
39.00	1199.0
40.00	1849.0
43.90	622.0
45.00	470.0
47.10	692.0
48.00	338.0
49.00	2475.0
50.10	11184.0
51.00	3922.0
52.00	119.0
55.00	178.0
56.00	882.0
57.00	1552.0
60.00	561.0
61.00	2869.0
62.00	2766.0
63.00	2190.0
64.10	217.0
65.10	329.0
67.00	216.0
68.00	6665.0
69.00	6510.0
69.90	428.0
71.90	347.0
73.00	3207.0
74.00	10536.0
75.00	31864.0
76.00	2652.0
77.00	511.0
77.90	275.0
78.90	1686.0
79.90	535.0
81.00	1850.0
81.80	404.0
87.00	2602.0
88.00	3051.0
91.00	249.0
92.00	1966.0
93.00	3220.0
94.00	7470.0
95.10	70672.0
96.00	5049.0
97.00	100.0
103.90	338.0
104.90	124.0
105.90	266.0
116.00	198.0
117.00	383.0
117.90	280.0
118.90	351.0
127.90	235.0
128.90	121.0
130.00	273.0
134.90	152.0
140.90	573.0
142.90	679.0
145.90	120.0
147.80	189.0
154.80	125.0
155.20	122.0
171.90	628.0
173.10	632.0
174.00	63104.0
175.00	4499.0
176.00	60456.0
177.00	3797.0

Instrument :
MSVOA_U
ClientSampleId :
BFB61

m/z	Abundance
36.00	526.0
37.00	2600.0
38.00	2754.0
39.00	1222.0
40.00	1708.0
43.90	584.0
45.00	491.0
47.00	667.0
48.00	372.0
49.00	2464.0
50.00	11563.0
51.00	4169.0
51.90	107.0
55.00	168.0
56.00	908.0
57.00	1703.0
60.10	612.0
61.00	3111.0
62.00	2880.0
63.00	2319.0
64.10	293.0
65.10	375.0
67.00	213.0
68.00	7122.0
69.00	6987.0
70.00	457.0
71.90	386.0
73.00	3456.0
74.00	11044.0
75.00	34000.0
76.00	2905.0
77.00	542.0
77.90	303.0
78.90	1916.0
79.90	569.0
80.90	1956.0
81.80	425.0
86.00	120.0
87.00	2999.0
88.00	3351.0
91.00	248.0
92.00	2050.0
93.00	3371.0
94.00	8186.0
95.10	75480.0
96.10	5441.0
97.00	151.0
104.00	334.0
105.00	118.0
105.90	217.0
115.90	216.0
117.00	395.0
117.90	249.0
118.90	352.0
127.90	244.0
128.90	117.0
130.00	290.0
134.90	157.0
136.70	115.0
137.20	103.0
140.90	559.0
142.90	693.0
145.90	127.0
147.90	238.0
154.80	139.0
155.10	131.0
156.90	113.0
171.90	739.0
173.00	644.0
174.00	68888.0
175.00	4801.0
176.00	65928.0
177.00	4214.0

Instrument :
MSVOA_U
ClientSampleId :
BFB61

m/z	Abundance
35.90	458.0
37.10	2546.0
38.00	2510.0
39.00	1039.0
40.00	1624.0
43.90	538.0
45.00	539.0
46.90	688.0
48.00	332.0
49.00	2284.0
50.00	10915.0
51.00	3894.0
51.90	114.0
55.00	171.0
56.00	889.0
57.00	1653.0
60.10	579.0
61.00	2945.0
62.00	2785.0
63.10	2201.0
64.10	316.0
65.00	379.0
67.10	181.0
68.00	6946.0
69.00	6596.0
70.00	433.0
71.90	433.0
73.00	3330.0
74.00	10688.0
75.00	33800.0
76.00	2750.0
77.00	514.0
77.90	288.0
78.90	1955.0
79.90	554.0
80.90	1869.0
81.90	392.0
86.00	154.0
87.00	3019.0
88.00	3137.0
88.80	136.0
91.00	217.0
92.00	1913.0
93.10	3263.0
94.10	8320.0
95.00	74160.0
96.10	4960.0
97.00	177.0
104.00	319.0
105.90	224.0
115.80	230.0
116.90	391.0
117.80	184.0
119.00	323.0
128.00	245.0
130.00	262.0
136.70	142.0
140.90	559.0
142.10	117.0
142.90	616.0
145.90	107.0
147.90	262.0
154.80	147.0
156.90	132.0
171.90	776.0
173.00	588.0
174.00	68336.0
175.00	4722.0
176.00	65168.0
177.00	4307.0
206.90	111.0

Instrument :
MSVOA_U
ClientSampleId :
BFB61