

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031221\
 Data File : VU042569.D
 Acq On : 11 Mar 2021 17:30
 Operator : SY/MD
 Sample : M1631-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B47

Quant Time: Mar 12 01:32:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 12 01:30:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	81267	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	78366	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	35532	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	25944	4.79	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.80%
7) Chloroethane-d5	1.919	69	22096	5.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.80%
11) 1,1-Dichloroethene-d2	2.572	63	45240	3.70	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.00%
20) 2-Butanone-d5	4.649	46	76895	43.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.46%
24) Chloroform-d	5.073	84	53053	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.713	65	35674	5.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.20%
32) Benzene-d6	5.736	84	98324	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	6.700	67	32801	5.29	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.80%
41) Toluene-d8	7.906	98	92970	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
43) trans-1,3-Dichloroprop...	8.189	79	12857	4.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.80%
46) 2-Hexanone-d5	8.646	63	63898	46.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.18%
57) 1,1,2,2-Tetrachloroeth...	10.764	84	27797	5.03	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.60%
64) 1,2-Dichlorobenzene-d4	12.202	152	35083	5.45	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031221\
Data File : VU042569.D
Acq On : 11 Mar 2021 17:30
Operator : SY/MD
Sample : M1631-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B47

Quant Time: Mar 12 01:32:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 12 01:30:46 2021
Response via : Initial Calibration

