

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031221\
 Data File : VU042599.D
 Acq On : 12 Mar 2021 08:54
 Operator : SY/MD
 Sample : VU0312WBL03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK25

Quant Time: Mar 12 11:39:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Mar 12 02:42:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.260	114	85599	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.427	117	86234	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.822	152	40277	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	23145	4.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.20%
7) Chloroethane-d5	1.919	69	19072	4.41	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.20%
11) 1,1-Dichloroethene-d2	2.575	63	39204	3.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.80%
20) 2-Butanone-d5	4.649	46	64256	34.29	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	68.58%
24) Chloroform-d	5.073	84	48453	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
26) 1,2-Dichloroethane-d4	5.713	65	33783	4.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.40%
32) Benzene-d6	5.739	84	86549	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.80%
36) 1,2-Dichloropropane-d6	6.700	67	29102	4.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	85.40%
41) Toluene-d8	7.906	98	87990	4.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.20%
43) trans-1,3-Dichloroprop...	8.192	79	10967	3.44	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.80%
46) 2-Hexanone-d5	8.649	63	55430	36.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.66%
57) 1,1,2,2-Tetrachloroeth...	10.767	84	24415	4.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.40%
64) 1,2-Dichlorobenzene-d4	12.201	152	31472	4.31	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.20%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031221\
Data File : VU042599.D
Acq On : 12 Mar 2021 08:54
Operator : SY/MD
Sample : VU0312WBL03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK25

Quant Time: Mar 12 11:39:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Mar 12 02:42:43 2021
Response via : Initial Calibration

