

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031224\
 Data File : VU058052.D
 Acq On : 12 Mar 2024 12:13
 Operator : MD/SY
 Sample : P1636-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 DCNS3

Quant Time: Mar 13 02:00:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 13 01:58:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	143118	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	136752	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	61270	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	37271	3.620	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	72.400%	
7) Chloroethane-d5	1.904	69	35643	4.211	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	84.200%	
11) 1,1-Dichloroethene-d2	2.560	65	17793	3.993	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	79.800%	
20) 2-Butanone-d5	4.611	46	93512	55.117	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	110.240%	
24) Chloroform-d	5.052	84	90636	4.654	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.000%	
26) 1,2-Dichloroethane-d4	5.695	65	43118	4.886	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.800%	
32) Benzene-d6	5.721	84	176195	4.421	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	88.400%	
36) 1,2-Dichloropropane-d6	6.682	67	57088	4.712	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	94.200%	
41) Toluene-d8	7.891	98	149223	4.133	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	82.600%	
43) trans-1,3-Dichloroprop...	8.177	79	17434	4.185	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.600%	
46) 2-Hexanone-d5	8.627	63	76647	54.796	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	109.600%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	37896	5.101	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	102.000%	
66) 1,2-Dichlorobenzene-d4	12.190	152	50472	5.002	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	100.000%	
Target Compounds						
16) Methylene chloride	3.036	84	3184	0.303	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031224\
 Data File : VU058052.D
 Acq On : 12 Mar 2024 12:13
 Operator : MD/SY
 Sample : P1636-16
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_U
ClientSampleId :
 DCNS3

Quant Time: Mar 13 02:00:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR022924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 13 01:58:39 2024
 Response via : Initial Calibration

