

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU031319\
 Data File : VU029996.D
 Acq On : 12 Mar 2019 03:16
 Operator : JC/SP
 Sample : K1796-19
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF5R1

Quant Time: Mar 12 08:51:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Mar 12 07:07:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	504588	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	488528	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.48	152	235382	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	174699	52.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.68%
7) Chloroethane-d5	1.68	69	141393	52.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	104.96%
11) 1,1-Dichloroethene-d2	2.27	63	234316	38.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.78%
21) 2-Butanone-d5	4.17	46	253698	101.45	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.45%
24) Chloroform-d	4.65	84	296068	49.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.90%
26) 1,2-Dichloroethane-d4	5.31	65	180166	52.09	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.18%
32) Benzene-d6	5.34	84	661330	52.59	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.18%
36) 1,2-Dichloropropane-d6	6.33	67	200359	52.06	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	104.12%
41) Toluene-d8	7.56	98	598031	50.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.58%
43) trans-1,3-Dichloropropene-	7.85	79	89665	48.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	96.18%
47) 2-Hexanone-d5	8.31	63	213708	99.24	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.24%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	291309	49.26	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.52%
64) 1,2-Dichlorobenzene-d4	11.86	152	241486	52.96	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.92%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU031319\
Data File : VU029996.D
Acq On : 12 Mar 2019 03:16
Operator : JC/SP
Sample : K1796-19
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF5R1

Quant Time: Mar 12 08:51:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Mar 12 07:07:04 2019
Response via : Initial Calibration

