

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031321\
 Data File : VU042647.D
 Acq On : 13 Mar 2021 06:49
 Operator : SY/MD
 Sample : M1676-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AE7

Quant Time: Mar 13 07:49:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Mar 13 06:39:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.257	114	75570	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.424	117	74686	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.819	152	32976	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	21892	4.35	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.00%
7) Chloroethane-d5	1.919	69	19712	5.17	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.40%
11) 1,1-Dichloroethene-d2	2.575	63	35664	3.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.60%
20) 2-Butanone-d5	4.649	46	89696	54.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.46%
24) Chloroform-d	5.073	84	48616	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.60%
26) 1,2-Dichloroethane-d4	5.713	65	33874	5.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.60%
32) Benzene-d6	5.736	84	87902	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.00%
36) 1,2-Dichloropropane-d6	6.697	67	29048	4.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.40%
41) Toluene-d8	7.906	98	83156	4.65	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.00%
43) trans-1,3-Dichloroprop...	8.186	79	10666	3.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.40%
46) 2-Hexanone-d5	8.646	63	61099	46.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.48%
57) 1,1,2,2-Tetrachloroeth...	10.765	84	25636	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	12.202	152	30817	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%
Target Compounds						
47) Tetrachloroethene	8.562	164	9881	1.87	ug/L	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031321\
Data File : VU042647.D
Acq On : 13 Mar 2021 06:49
Operator : SY/MD
Sample : M1676-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AE7

Quant Time: Mar 13 07:49:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SOMUTR030221WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Mar 13 06:39:26 2021
Response via : Initial Calibration

