

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU031420\
 Data File : VU037261.D
 Acq On : 14 Mar 2020 14:30
 Operator : JC/MD
 Sample : VU0314WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK23

Quant Time: Mar 16 05:26:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR031120WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Mar 16 05:16:55 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.28	114	124342	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.44	117	122437	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	54304	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	36946	4.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.40%
7) Chloroethane-d5	1.93	69	32627	5.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.20%
11) 1,1-Dichloroethene-d2	2.59	63	54913	3.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.20%
20) 2-Butanone-d5	4.68	46	110060	52.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.40%
24) Chloroform-d	5.11	84	77780	4.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.20%
26) 1,2-Dichloroethane-d4	5.74	65	48002	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.00%
32) Benzene-d6	5.76	84	145145	4.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.40%
36) 1,2-Dichloropropane-d6	6.73	67	47335	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	7.93	98	132293	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
43) trans-1,3-Dichloropropene-	8.21	79	19587	4.99	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	99.80%
46) 2-Hexanone-d5	8.66	63	92562	56.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	113.16%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	39641	4.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.00%
64) 1,2-Dichlorobenzene-d4	12.21	152	49528	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU031420\
Data File : VU037261.D
Acq On : 14 Mar 2020 14:30
Operator : JC/MD
Sample : VU0314WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK23

Quant Time: Mar 16 05:26:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR031120WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Mar 16 05:16:55 2020
Response via : Initial Calibration

