

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031422\
 Data File : VU047540.D
 Acq On : 14 Mar 2022 17:19
 Operator : SY/MD
 Sample : VU0314WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK079

Quant Time: Mar 15 05:45:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR031422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 15 05:45:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	151056	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	148615	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	75989	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	45015	4.557	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.200%
7) Chloroethane-d5	1.919	69	41926	4.676	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.600%
11) 1,1-Dichloroethene-d2	2.571	65	20759	4.481	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	89.600%
20) 2-Butanone-d5	4.645	46	146424	51.513	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.020%
24) Chloroform-d	5.070	84	99307	4.578	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.706	65	55458	4.718	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
32) Benzene-d6	5.729	84	198385	4.586	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
36) 1,2-Dichloropropane-d6	6.693	67	64915	4.771	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.899	98	177017	4.364	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
43) trans-1,3-Dichloroprop...	8.182	79	25962	4.539	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	8.639	63	145360	49.026	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.060%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	62815	4.623	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.400%
66) 1,2-Dichlorobenzene-d4	12.195	152	76292	4.805	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU031422\
Data File : VU047540.D
Acq On : 14 Mar 2022 17:19
Operator : SY/MD
Sample : VU0314WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK079

Quant Time: Mar 15 05:45:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR031422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Mar 15 05:45:04 2022
Response via : Initial Calibration

