

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
 Data File : VU030045.D
 Acq On : 13 Mar 2019 21:20
 Operator : JC/SP
 Sample : K1885-04
 Misc : 5.88g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BF188

Quant Time: Mar 14 04:36:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 14 03:50:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.88	114	555186	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	552095	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	296983	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	117962	32.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.24%
7) Chloroethane-d5	1.64	69	55221	18.63	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	37.26%#
11) 1,1-Dichloroethene-d2	2.23	63	167531	25.27	ug/L	-0.05
Spiked Amount	50.000	Range	60 - 125	Recovery	=	50.54%#
21) 2-Butanone-d5	4.19	46	228920	83.20	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.20%
24) Chloroform-d	4.64	84	247896	37.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.26%
26) 1,2-Dichloroethane-d4	5.30	65	154416	40.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.16%
32) Benzene-d6	5.33	84	527717	37.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.26%
36) 1,2-Dichloropropane-d6	6.32	67	171827	39.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	79.02%
41) Toluene-d8	7.56	98	482124	36.23	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	72.46%#
43) trans-1,3-Dichloropropene-	7.85	79	77159	36.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.24%
47) 2-Hexanone-d5	8.32	63	223207	91.71	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	91.71%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	305973	45.78	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.56%
64) 1,2-Dichlorobenzene-d4	11.86	152	236918	41.18	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.36%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU031519\
Data File : VU030045.D
Acq On : 13 Mar 2019 21:20
Operator : JC/SP
Sample : K1885-04
Misc : 5.88g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BF188

Quant Time: Mar 14 04:36:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM031319WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 14 03:50:43 2019
Response via : Initial Calibration

